

CDC Hormone Standardization Program (CDC HoSt)

Certified Estradiol Assays

From 2019 Q4 and forward

(UPDATED 04/2026)

CDC HoSt Programs started quarterly certification from November 2019 shipment.

See <https://www.cdc.gov/clinical-standardization-programs/media/pdfs/2024/04/CDC-Certified-Estradiol-Procedures-508.pdf> for previous list of certified assays.

- The following assays have successfully met the performance criteria of $\pm 12.5\%$ mean bias (for samples >20 pg/mL) and ± 2.5 pg/mL absolute bias (for samples ≤ 20 pg/mL) when compared to the CDC reference measurement procedure for estradiol for 80% of samples.
- It is not the intent of the CDC HoSt Program to certify each lot of reagents. Participants are awarded certificates for successfully meeting bias criteria using specific methods that consist of different reagent lots and calibrator lots.
- Analytical performance in CDC HoSt Program is assessed using human serum. The measurement procedures may have different accuracy and precision with other specimen types, such as plasma.
- Certification is valid for one quarter from the certification date. It is the responsibility of the participant to ensure that the results of the assay remain consistent, between lots, and over the measurement range reported.
- The analytical performance evaluation used in certification is for testing performed in patient care. Therefore, this certification does not imply suitability of a participant as a calibration laboratory or the procedure as a metrological reference measurement procedure.

Each table includes information about certified assays including participant name, measurement principle, assay identifier, assay measurement range, certification measurement range, certification date, individual samples pass rate, and contact information.

“Assay identifier” is an internal code used by the participant to represent the assay used for certification.

“Assay Measurement range” is the assays’ reported analytical measurement range (AMR) and is not the certification range.

“Certification Measurement Range” is the concentration range the of the samples used for HoSt certification.

“Certification date” includes historical certification information and gaps between years do not always indicate the assay’s failure to meet certification criteria.

“Individual samples pass rate” is the percentage of individual samples out of the 40 provided that met the certification criteria. This information was provided on the website starting June 2019.

Note: The $\pm 12.5\%$ mean bias used for certification can be considered the allowable calibration bias. Certification indicates that the assay is calibrated to meet those limits. Due to differences in test selectivity, measurements on individual samples can exceed the calibration bias. Therefore, the individual sample pass rate provides some information about the selectivity of a test that meets the calibration criteria.

Table 1: Currently Certified Assays including their certification history

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (pg/mL)	Certification Measurement Range (pg/mL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
ARUP Laboratories Salt Lake City, UT	LC/MS/MS†	Estradiol by TMS	2 - 1000	2.55 - 579	Q4 2025	90	Kayla West Kayla.west@aruplab.com 801-583-2787 x2893
				3.01 - 579	Q3 2025	92	
				3.01 - 579	Q2 2025	92	
				3.01 - 579	Q1 2025	100	
				2.55 - 579	Q4 2024	100	
				2.55 - 464	Q3 2024	100	
				2.55 - 464	Q2 2024	100	
				2.55 - 230	Q1 2024	100	
				3.05 - 230	Q4 2023	90	
				3.40 - 230	Q3 2023	88	
				2.70 - 230	Q2 2023	88	
				2.70 - 230	Q1 2023	85	
				2.50 - 230	Q4 2022	95	
				2.55 - 230	Q3 2022	90	
				2.55 - 230	Q2 2022	90	
				2.50 - 216	Q1 2022	98	
				2.70 - 230	Q4 2021	98	
				2.70 - 230	Q3 2021	100	
				3.01 - 230	Q2 2021	95	
				3.05 - 230	Q1 2021	95	
				2.55 - 230	Q4 2020	92	

Brigham Research Assay Core (BRAC) Laboratory at Harvard Medical School Boston, MA	LC/MS/MS ⁺	Serum Estradiol	1.00 - 500.00 (& higher than 500 pg/mL with dilution)	2.55 - 579	Q4 2025	95	Dr. Shalender Bhasin SBHASIN@PARTNERS.ORG (617)525-9040 Liming Peng Lpeng2@partners.org (617)525-9048
				2.55 - 230	Q3 2025	92	
				2.55 - 230	Q2 2025	95	
				2.55 - 464	Q1 2025	95	
				2.55 - 464	Q4 2024	80	
				2.70 - 464	Q3 2024	85	
				2.70 - 464	Q2 2024	85	
				2.55 - 230	Q1 2024	82	
				2.55 - 230	Q4 2023	92	
				2.55 - 230	Q3 2023	92	
				2.50 - 230	Q2 2023	92	
				2.50 - 230	Q1 2023	92	
				2.50 - 230	Q4 2022	95	
				2.50 - 230	Q3 2022	90	
				3.01 - 230	Q2 2022	90	
				2.55 - 230	Q1 2022	85	
				2.55 - 230	Q4 2021	88	
				2.55 - 216	Q3 2021	92	
				2.55 - 216	Q2 2021	92	
				2.70 - 216	Q1 2021	98	
				2.55 - 230	Q4 2020	98	
				2.55 - 230	Q3 2020	90	
				2.55 - 230	Q2 2020	95	
				2.55 - 230	Q1 2020	92	
				3.40 - 268	Q4 2019	92	
Centre Hospitalier Universitaire de Liège Sart-Tilman, Belgium	LC/MS/MS ⁺	25(OH)D3+25(OH)D2+C3-epimer+24,25(OH)2D in serum and plasma	5 - 1156	5.62 - 579	Q1 2025	82	Etienne Cavalier etienne.cavalier@chu.ulg.ac.be +3243667692
				5.62 - 579	Q4 2024	82	
				5.62 - 464	Q2 2024	80	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (pg/mL)	Certification Measurement Range (pg/mL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
				9.64 - 230 9.64 - 230 9.64 - 230 6.21 - 216	Q4 2023 Q1 2023 Q4 2022 Q2 2020	80 84 86 85	
Clinical Chemistry Branch CDC Atlanta, GA	LC/MS/MS ⁺	Total Estradiol in Serum (1036)	1.72 - 17,100	3.01 - 579 3.05 - 230 3.05 - 230 3.05 - 230 3.05 - 230 3.05 - 230 4.10 - 230 3.01 - 230 3.01 - 230 3.01 - 230 1.75 - 230 1.75 - 230 1.75 - 230 1.75 - 230 2.55 - 216	Q4 2025 Q1 2023 Q4 2022 Q3 2022 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	88 98 98 100 98 98 98 98 100 100 100 100 100 98 98	Lumi Duke, MS LDuke@cdc.gov (770)488-4126

Covance Central Laboratory Services Indianapolis, IN	LC/MS/MS+	Total Estradiol in Serum (E2)	2 - 4,000	2.70 - 579	Q4 2025	100	Cristina Hedin, MS Covance Central Laboratory Services Cristina.Hedin@covance.com 317-273-7842
				3.01 - 579	Q3 2025	100	
				3.05 - 579	Q2 2025	100	
				3.05 - 579	Q1 2025	100	
				3.40 - 579	Q4 2024	100	
				4.07 - 464	Q3 2024	98	
				2.50 - 464	Q2 2024	95	
				2.50 - 464	Q1 2024	92	
				2.50 - 230	Q4 2023	85	
				2.50 - 230	Q3 2023	87	
				3.05 - 230	Q2 2023	90	
				2.55 - 230	Q1 2023	92	
				2.55 - 230	Q4 2022	100	
				2.55 - 230	Q3 2022	100	
				2.55 - 230	Q2 2022	100	
				2.70 - 230	Q1 2022	100	
				2.70 - 230	Q4 2021	100	
				2.50 - 230	Q3 2021	100	
				2.50 - 230	Q2 2021	100	
				2.50 - 230	Q1 2021	100	
				2.50 - 230	Q4 2020	100	
				2.70 - 230	Q3 2020	100	
				2.55 - 268	Q2 2020	100	
				2.55 - 268	Q1 2020	100	
				2.55 - 268	Q4 2019	90	
LabCorp Burlington, NC	LC/MS/MS+	Estradiol in Serum	2.5 - 5000	4.46 - 382	Q3 2025	82	Majid Moridani moridam@labcorp.com 336-436-3102
				4.07 - 230	Q2 2021	90	
				2.55 - 230	Q1 2021	90	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (pg/mL)	Certification Measurement Range (pg/mL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
				2.55 - 230	Q4 2020	90	
				2.55 - 230	Q3 2020	88	
				2.55 - 230	Q2 2020	85	

LabCorp Calabasas Hills, CA	LC/MS/MS ⁺	#500108 Estradiol, LC/MS (Endocrine Sciences)	1 - 500 (1 to 5,000 with validated dilution)	2.70 - 464	Q4 2025	90	Majid Moridani moridam@labcorp.com 336-436-3102 Dr. Brett Holmquist holmquib@labcorp.com (818) 867-1362 Dr. Kelly Chun chunk@labcorp.com (818) 867-1358
				3.01 - 464	Q3 2025	92	
				3.01 - 464	Q2 2025	82	
				3.01 - 382	Q1 2025	87	
				3.01 - 382	Q4 2024	87	
				3.01 - 382	Q3 2024	87	
				2.55 - 382	Q2 2024	97	
				2.55 - 230	Q1 2024	92	
				9.64 - 230	Q4 2023	92	
				2.50 - 230	Q3 2023	92	
				2.50 - 230	Q2 2023	92	
				2.50 - 230	Q1 2023	95	
				2.50 - 230	Q4 2022	90	
				2.70 - 230	Q3 2022	88	
				2.70 - 230	Q2 2022	85	
				2.70 - 230	Q1 2022	88	
				2.55 - 230	Q4 2021	95	
				2.55 - 230	Q3 2021	98	
				2.55 - 230	Q2 2021	100	
				2.55 - 230	Q1 2021	98	
				2.55 - 230	Q4 2020	98	
				2.55 - 230	Q3 2020	90	
				2.55 - 230	Q2 2020	88	
				2.55 - 268	Q1 2020	90	
				3.40 - 268	Q4 2019	90	
LabCorp Spokane, WA	LC/MS/MS ⁺	ESTRADIOL (LCMSMS)	2.5 - 625 (2.5 - 5000 with validated dilution)	4.46 - 579	Q4 2025	88	Carissa Schmitz MLS(ASCP)CM Schmic4@LabCrop.com (509) 755-8358
				4.46 - 579	Q3 2025	92	
				4.46 - 579	Q2 2025	88	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (pg/mL)	Certification Measurement Range (pg/mL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
				4.46 - 579	Q1 2025	88	
				4.07 - 579	Q4 2024	82	
				4.46 - 230	Q4 2023	80	
				3.95 - 230	Q3 2023	95	
				3.95 - 230	Q2 2023	92	
				3.95 - 230	Q1 2023	90	
				3.95 - 230	Q4 2022	88	
				4.70 - 230	Q3 2022	85	
				4.07 - 230	Q2 2022	87	
				4.07 - 230	Q1 2022	92	
				4.07 - 230	Q4 2021	90	
				3.95 - 230	Q3 2021	88	
				3.95 - 230	Q2 2021	85	
				3.95 - 230	Q1 2021	85	
				3.95 - 230	Q4 2020	90	
				4.97 - 230	Q3 2020	88	
				6.21 - 230	Q2 2020	95	
				6.21 - 230	Q1 2020	98	
				5.90 - 216	Q4 2019	88	

Mayo Clinic Rochester, MN	LC/MS/MS ⁺	Estradiol	10 - 600	11.7 - 230	Q4 2025	100	Sue Reicks reicks.sue@mayo.edu
				11.7 - 230	Q3 2025	100	
				11.7 - 464	Q2 2025	100	
				11.7 - 464	Q1 2025	100	
				12.3 - 464	Q4 2024	95	
				11.7 - 464	Q3 2024	95	
				11.7 - 464	Q2 2024	95	
				11.7 - 216	Q1 2024	95	
				11.7 - 230	Q4 2023	100	
				11.7 - 230	Q3 2023	98	
				11.7 - 230	Q2 2023	95	
				11.7 - 230	Q1 2023	95	
				11.7 - 230	Q4 2022	95	
				11.7 - 230	Q3 2022	92	
				11.7 - 230	Q2 2022	95	
				11.7 - 230	Q1 2022	100	
				11.4 - 230	Q4 2021	100	
				11.4 - 230	Q3 2021	100	
				11.4 - 230	Q2 2021	100	
				11.4 - 230	Q1 2021	100	
				11.7 - 230	Q4 2020	100	
				11.8 - 230	Q3 2020	95	
				17.1 - 230	Q2 2020	95	
				17.1 - 216	Q1 2020	95	
				16.6 - 230	Q4 2019	92	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (pg/mL)	Certification Measurement Range (pg/mL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Roche Diagnostics GmbH Penzberg, Germany	LC/MS/MS ⁺	Total Estradiol in Serum and Plasma	5 - 5000	9.64 - 579	Q4 2024	100	Judith Taibon judith.taibon@roche.com 0049 8856 60 12941
				9.64 - 382	Q3 2024	100	
				9.64 - 382	Q2 2024	90	
				9.64 - 230	Q1 2024	90	
				9.64 - 230	Q4 2023	90	
				9.64 - 230	Q3 2023	90	
				9.64 - 230	Q2 2023	100	
				9.64 - 230	Q1 2023	100	
				11.4 - 230	Q4 2022	100	
				9.64 - 230	Q3 2022	100	
				9.64 - 230	Q2 2022	92	
				9.34 - 216	Q1 2022	92	
				9.34 - 230	Q4 2021	92	
				9.34 - 230	Q3 2021	92	
				9.34 - 230	Q2 2021	95	
				9.64 - 230	Q1 2021	92	
				9.64 - 230	Q4 2020	92	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (pg/mL)	Certification Measurement Range (pg/mL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Roche Diagnostics GmbH Munich, Germany	LC/MS/MS†	Estradiol	0.1 - 400	2.55 - 579	Q4 2025	100	Dr. Stefanie Grimm muc.testsidestudy@roche.com
				3.05 - 579	Q3 2025	100	
				3.01 - 579	Q2 2025	100	
				3.01 - 579	Q1 2025	100	
				2.50 - 579	Q4 2024	100	
				2.50 - 464	Q3 2024	100	
				2.50 - 464	Q2 2024	100	
				2.50 - 216	Q1 2024	100	
				2.50 - 230	Q4 2023	100	

† LC/MS/MS – Liquid Chromatography Tandem Mass Spectrometry